



America


CERTIFICATE

No. U8 017955 0148 Rev. 00

Holder of Certificate: Omicron electronics GmbH

Oberes Ried 1
6833 Klaus
AUSTRIA

Certification Mark:



C US

Product: Electronic measuring equipment

Model(s): ARCO 400
CMC430

Parameters:	ARCO 400	CMC430	
	Rated voltage:	100-240 V	100-240 V
	Rated frequency:	50/60 Hz	50/60 Hz
	Rated current:	10-8 A	
	Rated power:		1000 W
Protection class:	I	I	

Tested according to:

UL 61010-1:2012/R:2019-07
UL 61010-2-030:2018
CSA C22.2 No. 61010-1:2012/A1:2018-11
CSA C22.2 No. 61010-2-030:2018

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713266272

Date, 2022-08-10

Siemon

(Thorsten Siemon)